


SRV® 3 TEST SYSTEM



HW Config	Oscillation	Printout Date	25.01.2017 10:32:59
Test No	m02225	Measurement Date	25.01.2017 07:39:26
Operator	Bogusch	Test Mode	hoch belasteter Motor.opr
Upper Specimen	Kugel 10mm	Type of Test	Laborauftrag
Interfacial Media	Öl	Note	Handelsübliches 15W40
Lower Specimen	Disc 24*7,9 100CR79		

Comment



Friction coeff Frequency [Hz] Load [N] Stroke [µm] Temp. Block [°C]	Results: f min 0,095 f max 2,559 f 15 0,111 f 30 0,104 f 90 0,098 f 120 0,098 Wk [mm] 0,55 Wear (man) 0,8
Optimol Instruments Prüftechnik GmbH, München/Germa	