


SRV® 3 TEST SYSTEM



HW Config	Oscillation	Printout Date	25.01.2017 13:07:05
Test No	m02226	Measurement Date	25.01.2017 10:43:18
Operator	Dangel	Test Mode	hoch belasteter Motor.opr
Upper Specimen	Kugel 10mm	Type of Test	Laborauftrag
Interfacial Media	Öl + Additiv	Note	Handelsübliches 15W40+ 5% MoS2 Suspensio
Lower Specimen	Disc 24*7,9 100CR79		

Comment



- Friction coeff
- Frequency [Hz]
- Load [N]
- Stroke [µm]
- Temp. Block [°C]

Results:

f min	f max	f 15	f 30	f 90	f 120	Wk [mm]	Wear (man)
0,096	1,146	0,109	0,101	0,098	0,099	0,52	0,7

Optimol Instruments Prüftechnik GmbH, München/Germa